



IMDRF
International Medical Device
Regulators Forum

EU2023
EUROPEAN UNION
Chair



European
Commission



European
Union

Healthcare professional engagement in Post-market surveillance of Korea

You Kyoung Lee,

Dept. of Laboratory Medicine & Medical Device Clinical Research Center (MDRC),

SoonChunHyang University Bucheon Hospital

2023-03-27



OVERVIEW

MDSIM consortium in Korea	3
MDAE & reportable MDAE	4
MDAE reporting in Korea	5
MDSIM; Healthcare professional engagement	6
What challenges we recognized	8
Perspective diversity	9
Business-cultural aspect	11
What efforts are we making	12
Engage in MDAE reporting, Why?	13
Tasks in MDAE reporting; health profession's viewpoint	14
What efforts are we making	15



IMDRF
International Medical Device
Regulators Forum

EU2023
EUROPEAN UNION
Chair

Healthcare professional engagement in Post-market surveillance of Korea

Medical device adverse event reporting by healthcare institutions in Korea, 12 years experience


MDAE vs reportable MDAE

Medical Device Adverse Event (MDAE)¹

- an unexpected event that occurs during or result from 'patient use' of a medical device

Reportable MDAE²



- An event has occurred
- The device is associated with the event
- The event led to one of the following outcomes
 - Death or Serious injury of a patient, user or other person
 - No death or serious injury occurred but the event might lead to death or serious injury



References:

1. C Yoon et al. *Differences in Perspectives of Medical Device Adverse Events: Observational Results in Training Program Using Virtual Cases*. J Korean Med Sci. 2019 Oct 14;34(39):e255
2. GHTF/SG2/N21R8:1999 (GHTF/FD:99-7)

MDAE reporting in Korea



S Choi et al. *The establishment of the Korean medical device safety information monitoring center: Reviewing ten years of experience.* Health policy 125 (2021) 941-946

Abbreviations: AER; Adverse Event Reporting, MFDS; Ministry of Food and Drug Safety, NIDS; National Institute of Medical Device Safety Information

MDSIM: Healthcare professional engagement

Medical Device Safety Information Monitoring Center (MDSIM) Pilot (2010)

- Two general hospitals
 - 60 (include 27 serious) cases collected through a medical record review
- Lessons learned
 - Need to collect all MDAEs from serious to mild
 - Need to improve awareness to the MDAE and AER in the healthcare practice field




Fig. 1. Numbers of medical device adverse event reports by reporting year and source in Korea.
Note: MDSIMC = Medical Device Safety Information Monitoring Center.

S Choi et al. *The establishment of the Korean medical device safety information monitoring center: Reviewing ten years of experience.* Health policy 125 (2021) 941


MDSIM; Brief history

The settle-down period (2012-2017)

- Establishing MDSIM consortium
 - 6 designated certified tertiary hospitals (regional center)
- Implement 'AE Review Committee' in each regional center
- Implement 'adverse event terminology system'

The challenge period (2018-present)

- Expanding MDSIM
 - 17 regional centers & 123 affiliated healthcare institutions
- MFDS established 'AE expert committee' (2018)






IMDRF
International Medical Device
Regulators Forum

EU2023
EUROPEAN UNION
Chair

Healthcare professional engagement in Post-market surveillance of Korea

What challenges we recognized

Perspective diversity



Which is the best fit moment to define as 'patient use'?

Fig. 3. Stakeholder's response rate for phase of using medical device for patient

Lee YJ, et al. Perspective Diversity of Domestic Stakeholders on Medical Device Adverse Event Reporting. *Journal of Biomedical Engineering Research* 40:171;2019





Fig. 2. Differences in results of group discussion while applying the same guidelines for virtual cases. AE = adverse event, NCAR = National Competent Authority Report.

C Yoon et al. Differences in Perspectives of Medical Device Adverse Events: Observational Results in Training Program Using Virtual Cases. *J Korean Med Sci.* 34(39):e255;2019

Business-cultural aspect

Lack of awareness

- It happens frequently and quite natural!
- Necessary to report?

Cultural immaturity

- Fear of blame
- Too busy! Why me?




Figure 1 Factors influencing adverse medical device event (AMDE) reporting.

Gagliardi AR, et al. Factors influencing the reporting of adverse medical device events: qualitative interviews with physicians about higher risk implantable devices. *BMJ Qual Saf* 27:190–198;2018



IMDRF
International Medical Device
Regulators Forum

EU2023
EUROPEAN UNION
Chair

Healthcare professional engagement in Post-market surveillance of Korea

What efforts are we making

Engage in MDAE reporting, Why?

Quality improvement in healthcare practice


- Improve patient outcome
- Information fed into the quality management system of a healthcare institution

Contribute to upgrading medical device

- Identify clinical unmet needs
- Risk management by manufacturer

Tasks in MDAE reporting; health profession's viewpoint

- Implement 'Just Culture'
- Collecting MDAEs
 - Should not be limited to the reportable MDAEs
- Know the approximate event rate?
 - Need to build a system that can identify MDAEs against usage in healthcare practice (real-world data)



What efforts are we making

Education & training program

- MDSIM consortium
 - Regular case review meetings with IMDRF code application training
- APEC Center of Excellence programs

Research activities

- Feasibility study to construct a big-data system for MDV
- Pilot study to capture UDI from MD during the healthcare procedure




Figure 6. Ceramic femoral head hierarchy was applied to Stryker's products registered with national insurance EDI codes E1011031 and E1011231. EDI electronic data interchange, SNOMED systematized nomenclature of medicine clinical terms, UDI unique device identifier, DI device identifier.

S Choi, et al. Preliminary feasibility assessment of CDM-based active surveillance using current status of medical device data in medical records and OMOP-CDM. *Sci Rep* 11:24070;2021



IMDRF
International Medical Device
Regulators Forum

EU2023
EUROPEAN UNION
Chair

THANK YOU / QUESTIONS

You Kyoung Lee, Director of SCH APEC Center of Excellence (medical device)

youklee@gmail.com or cecilia@schmc.ac.kr

Disclaimer

This document was produced by the International Medical Device Regulators Forum. There are no restrictions on the reproduction or use of this document; however, incorporation of this document, in part or in whole, into another document, or its translation into languages other than English, does not convey or represent an endorsement of any kind by the International Medical Device Regulators Forum.

Copyright 2021 by the International Medical Device Regulators Forum.

